

Watson-Marlow Ltd
 Telephone +44 (0) 1326 370370
 Fax +44 (0) 1326 376009
 E-mail: info@wmpg.co.uk
www.wmpg.com

A Spirax Sarco Engineering Company



The information contained in this document is believed to be correct, but the Watson-Marlow Pumps Group accepts no liability for any errors it contains, and reserves the right to alter specifications without notice.


WARNING
 This product is not designed for use in, and should not be used for, patient-connected applications.

PB0388/10

	120S/DV 1-200rpm 120S/R 1-32rpm 120S/D1 1-200rpm 120S/DM2 1-100rpm 120S/DM3 1-100rpm		+5C, +41F
	<60dBA@1m		+40C, +104F
	120S/DV 1.5kg 120S/R 1.55kg 120S/D1 1.65kg 120S/DM2 1.7kg 120S/DM3 1.7kg		-25C, -13F
			+65C, +149F



120S/DV, 120S/R, 120S/D1,
 120S/DM2, 120S/DM3



David Cole, Managing director



2006/42/EC, 2004/108/EC,
 BS EN60529 (IP31), BS EN 61010-1,
 UL 61010-1,
 CAN/CSA-C22.2 No 61010-1



